


Lovibond® Colour Measurement Tintometer® Group



Gardner

Part Number: 342000



Conformance filters are assigned a nominal value and the company always endeavors to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Gardner

A set of two glass conformance filters for checking the 3000 Comparator in Gardner scale

Technical Data

Glass Standard Values	Approximate Values
	- 3
	- 17
Filters	2

Delivery Scope

- Gardner 3 filter
- Gardner 17 filter
- Certificate of conformity