


Optional Accessories SERIES 544 — Laser Scan Micrometer (Display Unit)

Digimatic code output unit

- 2-channel Digimatic code output
- In simultaneous measurement, measurement data are output as follows:
Program No. 0 to No. 4 in OUTPUT-1
Program No. 5 to No. 9 in OUTPUT-2
(10 programs operated)
- 10 pin MIL type connector.
- Output cable is not supplied.
Connecting cable (optional) 1 m (936937)
- Available for **LSM-6200 (544-071, 544-072)** and **LSM-6902H (544-498-1, 544-499-1)**.

Note 1: Output is 6 digits of measurement data.


Note 2: Displaying 6th and 7th digit after the decimal point is not supported.



SPECIFICATIONS


Order No.	02AGC840
-----------	----------

XY Measurement



(X-Y): runout
(X+Y)/2: average
Note: XY requires 10 mm-interval.

Parallel Measurement




Dual connection unit

- Enables second unit connection to **LSM-6200 (544-071, 544-072)**. (both units must be the same model)

Note: Cannot be used for **LSM-6902H (544-498-1, 544-499-1)**.

- Depending on the layout of the two measuring units, large-diameter measurement, XY measurement, and parallel measurement are possible.
- Both of the measuring units and display units can be simultaneously operated.



SPECIFICATIONS


Order No.	02AGP150
-----------	----------

2nd I/O analog I/F

- I/O, analog output.
- Simultaneous measurement is supported by two pairs of GO/NG judgment outputs.
- Available for **LSM-6200 (544-071, 544-072)** and **LSM-6902H (544-498-1, 544-499-1)**.

SPECIFICATIONS

Order No.	02AGC880
Standard Accessories	Connector (DDK) 57-30360 (214188)



Cable for BCD and 2nd I/O simultaneous mount


- Both BCD (02AGC910) and 2nd I/O analog I/F (02AGC880) can be mounted on **LSM-6200 (544-071, 544-072)** and **LSM-6902H (544-498-1, 544-499-1)** using this cable.

Note: If using this cable, the dual connection unit (02AGP150) cannot be used.

SPECIFICATIONS

Order No.	02AGE060
-----------	----------

DIMENSIONS



Unit: mm