



UKCA – Declaration of Conformity

We declare that the product **Temperature Calibrator**

series **TP37700E2, TP37700E2i**

manufactured by **SIKA Dr. Siebert & Kühn GmbH & Co. KG**

comply with

Electrical Equipment (Safety) Regulations 2016

Electromagnetic Compatibility Regulations 2016

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

The devices comply with following technical specifications

BS EN 61010-1:2010+ A1:2019	Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements
BS EN IEC 61010-2- 010:2020	Safety requirements for electrical equipment for measurement, control and laboratory use. Particular requirements for laboratory equipment for the heating of materials
BS EN 61326-1:2013-02	Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements
BS EN IEC 63000:2018	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

Kaufungen, 08.07.2021

i.V. 

Dipl.- Ing. K. Ulloth
(Manager CE- Coordination and Safety Supervisor)
(CE- Koordinator und Produktsicherheitsbeauftragter)