

# Solution Expiration Dates

## A GUIDE FOR ENVIRONMENTAL PROFESSIONALS



**NOTE:** Expiration dates are clearly labeled on the bottle. Expiration times for unopened / unmixed solutions are from date of manufacture, not date of purchase. TruLine solutions are for the TruLine laboratory ISEs.

CONDUCTIVITY	
Unopened - Glass Quarts	1 Year
Unopened - Plastic Pints	18 Months
Opened	1 Month

TruLine ISE CALIBRATION, ISA & FILLING SOLUTIONS	
Unopened	15 Months
Opened	6 Months

TURBIDITY	
Unopened	1 Year
Opened	6 Months

ZOBELL	
Unmixed	5 Years
Mixed	6 Months

AMMONIUM	
Unopened	18 Months
Opened	6 Months

pH	
Unopened	2 Years
Opened	2 Months

NITRATE	
Unopened	1 Year
Opened	6 Months

CONFIDENCE SOLUTION	
Unopened	2 Years
Opened	3 Months

O <sub>2</sub> PROBE ELECTROLYTE SOLUTION	
Unmixed Crystals	No Stated Date
Mixed Solution	1 Year