Ergotron Carts

Cleaning recommendations





Ergotron carts deliver more than compact, mobile data entry at the point of care. Unparalleled features and superior construction help increase accuracy and efficiency—now and when technologies change. StyleView® and Neo-Flex® workstations are an economical investment built to exceed industry expectations.

All plastic and metal surfaces, whether painted, polished or unfinished will withstand cleaning by most commonly used hospital-grade, diluted, non-abrasive, non-corrosive solutions such as quaternary ammonia compounds, ammonia enzyme cleaners, hydrogen peroxide, bleach or alcohol solutions.


- Pen and permanent and dry erase markers can be removed with 91% isopropyl alcohol and a soft cloth.
- lodine stains can be removed with commonly used cleaners and a soft cloth.
- Never use steel wool or other abrasive materials that will damage the surface finish.
- Do not use strong solvents such as trichloroethylene and acetone. These solvents will damage the surface finish.

It is recommended that any cleaning solution be tested on a small, inconspicuous area to ensure surface is not harmed.

Recommended cleaning steps

The following procedures are not guaranteed to control infection. The hospital's infection control administrator or epidemiologist should be consulted regarding cleaning procedures and processes.





Do not use flammable cleaners on cart surfaces due to close proximity of electrical power and equipment.

© 2009 Ergotron, Inc. rev. 02/10/2015 Content is subject to change without notification

Corporate Headquarters EMEA Sales APAC Sales St. Paul, MN USA Amersfoort, The Netherlands Tokyo, Japan (800) 888-8458 +31 33 45 45 600 www.ergotron.com apaccustomerservice@ergotron.com

Worldwide 0EM Sales www.ergotron.com info.oem@ergotron.com



Americas Sales and