

Keysight 8478B Option H27

Product Note

The Keysight 8478B Option H27 is calibrated at the standards lab at 9 MHz.

For information concerning the operation and connections, reference the standard Keysight 8478B document.

In all other respects this instrument is identical to the standard instrument specifications.

Inspect the shipping container. If the container or packing material is damaged, it should be kept until the contents of the shipment have been checked mechanically and electrically. If there is physical damage please notify the nearest Keysight Technologies office. Keep the damaged shipping materials (if any) for inspection by the carrier and a Keysight Technologies representative.

To contact Keysight for sales and technical support, refer to the Keysight worldwide website at: www.keysight.com/find/assist.



This information is subject to change without notice.

© Keysight Technologies 2005 - 2019
Edition 3, November 15, 2019

Printed in Malaysia



08478-90120

www.keysight.com