


P1053 500 MHz 10:1 oscilloscope probe

User's Guide



EC declaration of conformity

Pico Technology declares that the following products comply with the requirements of the specified Directives and Standards as listed below. Technical documentation required to demonstrate compliance to the standards is available for inspection by the relevant enforcement authorities. Products carry the CE mark.

Products covered by this declaration:

P1053 500 MHz 10:1 oscilloscope probe

EU Directives covered by this declaration:

2014/35/EU	<i>Low Voltage Equipment Directive</i>
2012/19/EU	<i>Waste Electrical and Electronic Equipment</i>
2011/65/EU	<i>Restriction of use of certain Hazardous Substances</i>

The basis on which conformity is being declared:

EN61010-031:2015	<i>Safety requirements for hand-held probe assemblies for electrical measurement and test.</i>
RoHS and WEEE	<i>Manufacturer's analysis of the raw materials used in the manufacture of the above products.</i>

Your help and efforts are required to protect and keep our environment clean. Therefore either return this product at the end of life to the manufacturer or ensure WEEE compliant collection and treatment yourself. Do not dispose of as unsorted municipal waste.



Warranty

Pico Technology warrants this oscilloscope accessory for normal use and operation within specifications for a period of two years from date of shipment and will repair or replace any defective product which was not damaged by negligence, misuse, improper installation, accident or unauthorized repair or modification by the buyer. This warranty is applicable only to defects due to material or workmanship. Pico Technology disclaims any other implied warranties of merchantability or fitness for a particular purpose. Pico Technology will not be liable for any indirect, special, incidental, or consequential damages (including damages for loss of profits, loss of business, loss of use or data, interruption of business and the like), even if Pico Technology has been advised of the possibility of such damages arising from any defect or error in this manual or product.

Safety

To prevent possible electrical shock, fire, personal injury, or damage to the product, carefully read this safety information before attempting to install or use the product. In addition, follow all generally accepted safety practices and procedures for working with and near electricity.

The product has been designed and tested in accordance with the European standard publication EN 61010-031:2015, and left the factory in a safe condition.

The following safety descriptions are found throughout this guide:

A **WARNING** identifies conditions or practices that could result in injury or death.

A **CAUTION** identifies conditions or practices that could result in damage to the product or equipment to which it is connected.

Symbols

These safety and electrical symbols may appear on the product or in this guide:

Symbol	Description	
	Earth (ground) terminal	Terminal can be used to make a measurement ground connection. The terminal is NOT a Safety or protective Earth.
	Possibility of electric shock	
	Caution	Appearance on the product indicates a need to read these safety and operation instructions.
	Do not dispose of this product as unsorted municipal waste.	

WARNING

To prevent injury or death only qualified personnel should use this product, only as instructed and with only accessories supplied or recommended. Protection provided by the product may be impaired if used in a manner not specified by the manufacturer.

Maximum input voltage

The table and frequency derating plot below indicate the maximum input voltage for these probes. This is the maximum voltage that can be safely and accurately measured using the probes. The maximum input voltage depends on the signal frequency, the measurement category and also on the instrument with which the probe is used.

WARNING

To prevent electric shock, do not attempt to connect voltages outside the probe's voltage rating, or above the oscilloscope's maximum input voltage multiplied by the attenuation ratio of the probe, whichever is lower. When using accessories with the probe, the lower rating / measurement category always applies to both probe and accessories connected to it. If an accessory is not marked with a voltage rating on either the connector, cable or body, or if a protective finger guard is removed, then do not exceed the EN61010 "hazardous live" limits overleaf.


Model	Instrument	Maximum input voltage
P1053	In combination with PicoScope 3000E*	±200 V DC+AC peak (not in CAT II, III or IV)
	Maximum for probes with any instrument	300 V RMS (CAT II)

* When used with the PicoScope 3000E, maximum input voltage to the probe is limited to ten times the instrument's maximum measuring range of ±20 V.

⚠ WARNING

To avoid overloading the probe, note that its maximum input voltage rating decreases as the frequency of the applied signal increases.

Typical voltage derating P1053
No Measurement Category



⚠ WARNING

To prevent electric shock, take all necessary safety precautions when working on equipment where hazardous live voltages may be present.

Signals exceeding the voltage limits in the table below are defined as "hazardous live" by EN 61010

Signal voltage limits of EN 61010-031:2015		
±60 V DC	30 V AC RMS	±42.4 V pk max.

⚠ WARNING

To prevent electric shock caused by accidental contact between a hazardous live voltage and the probe ground input, do not use the probe tip ground spring or ground blade when working around hazardous live voltages. Where the risk of contact between a hazardous live voltage and the probe ground cannot be avoided, a differential isolating probe such as the TA041 listed on the Pico website must be used instead.

⚠ WARNING

To prevent injury or death, do not use this probe for measurements on mains installations in measurement categories III and IV.

⚠ WARNING

To prevent injury or death, do not use the product or an accessory if it appears to be damaged in any way, and stop use immediately if you are concerned by any abnormal operations.

⚠ CAUTION

Exceeding the voltage rating of any cable, connector or accessory can cause permanent damage to the oscilloscope and other connected equipment.

Grounding

⚠ WARNING

Never connect the ground input to, or allow it to touch, any electrical potential other than ground. To prevent personal injury or death, use a voltmeter to check that there is no significant AC or DC voltage between the probe ground and the point to which you intend to connect it.

⚠ CAUTION

Applying a voltage to the ground input is likely to cause permanent damage to the probe or other connected equipment.

It is good practice to connect the probe output to the measurement instrument and the ground lead to earth ground before connecting the probe to the circuit under test. Disconnect the probe input and the probe ground lead from the circuit under test before disconnecting the probe from the measurement instrument.

External connections

CAUTION

Take care to avoid mechanical stress or tight bend radii for all connected leads, including all coaxial leads and connectors. Mishandling will cause deformation and will degrade performance and measurement accuracy.

Environment

WARNING

To prevent injury or death, do not use in wet or damp conditions or near explosive gas or vapor.

CAUTION

To prevent damage, always store your probe in appropriate environments.

	Storage	Operating
Temperature	-40 to 60 °C	-20 to 50 °C
Max. humidity (non-condensing)	80%	80% to 30 °C, decreasing linearly to 40% at 50 °C
Max. altitude	15 000 m	2 000 m
Pollution degree	2 As defined in IEC 61010-031. Only non-conductive pollution. Occasionally, however, a temporary conductivity caused by condensation must be accepted.	

Care of the product

The probe contains no user-serviceable parts. Repair, servicing and calibration require specialized test equipment and must only be performed by Pico Technology or an approved service provider. There may be a charge for these services unless covered by the Pico two-year warranty.

Inspect the instrument and all probes, connectors, cables and accessories before use for signs of damage.

WARNING

To prevent electric shock do not tamper with or disassemble the probe, case parts, connectors or accessories.

CAUTION

When cleaning the product, use a soft cloth moistened, if necessary, with either distilled water or isopropyl alcohol. Before use allow the probe to dry completely. To prevent electric shock, do not allow liquids to enter the probe casing, as this will compromise the electronics or insulation inside.

CAUTION

To avoid injury, handle with care especially when fitted with the extra thin and sharp spring contact tip.


About the probes

Pico Technology's P1053 probe is a high-performance oscilloscope probe with a 10:1 attenuation ratio and a 500 MHz bandwidth. This compact probe is ideal for both field service and development environments, offering excellent visibility on dense SMT circuits due to their slim 3.5 mm tip housing diameter.

Here's a closer look at their key features:

- **Wide bandwidth** 500 MHz for capturing high-frequency signals.
- **High voltage rating** 300 V CAT II for safe operation in various environments.
- **Low input capacitance** Only 9 pF minimizes loading on the signal source, particularly important for fast rise time signals.
- **Reduced inductance** Short ground leads minimize inductance and resonances for improved signal fidelity.
- **Spring-loaded tips** Ensures secure connection to test points.
- **Variety of accessories** Includes various probing attachments for adaptability.

Overall, the P1053 probe from Pico Technology offers a compelling combination of performance, size, and versatility for oscilloscope users working on demanding applications.



Specifications

This User's Guide supersedes all previously published material. Specifications that are not marked as guaranteed are published as general information to the user. The instrument should have warmed up for at least 20 minutes and the environmental conditions must not exceed the specified limits of the probe. Note that specifications are subject to change without notice.

Electrical specifications

	P1053	
Attenuation ratio*	10:1	±2% at DC
Voltage coefficient	0.00025%/V	at DC
Probe bandwidth	500 MHz	−3 dB
Probe risetime	700 ps	10% to 90%, typical

* Connected to oscilloscope with an input impedance of 1 MΩ ± 1%.


Electrical characteristics

	P1053	
Input resistance (system)	10 MΩ	±2%
Input capacitance (system)	9 pF	typical
Compensation range	9 to 54 pF	typical
Input coupling of the measuring instrument	1 MΩ AC / DC	

Input impedance

 **CAUTION**

To avoid overloading the circuit under test, note that the input impedance of the probe decreases as the frequency of the applied signal increases.





Mechanical characteristics

Weight (probe only)	40 g
Cable length	1.2 m
Probe tip diameter	3.5 mm


Frequency compensation

Before taking any measurements using the probe, first check its compensation and adjust it to match the channel inputs.

Connect the probe to a 2 V pk-pk, 1 kHz square wave source. Most PicoScope oscilloscopes have a signal generator output marked GEN or AWG, or a probe CAL pin, which you can configure to generate such a signal. Adjust the LF trimmer until you see a flat-top square wave on the display:



Probe accessories



For further information on available accessories please refer to the accessory kits listed on page 10.

Probe accessories

You can buy a range of kits containing accessories and spare parts for the P1053 probe. Contents of the standard product packs and optional kits are listed below.

	TA561 - P1053 Single Pack 500 MHz	TA562 - P1053 Dual Pack 500 MHz	TA563 - BNC adaptor 3.5 mm	TA564 - Replacement Tips Rigid	TA565 - Replacement Tips Sprung
Probe 500 MHz	1	2			
BNC adaptor			1		
Protection cap	1*	2*			
Set coding rings 4x2 colors	1	1			
Trimmer tool	1	1			
Ground lead 17 cm	1	2			
Rigid ground spring 3.5 mm	1	2			
Rigid tips 0.4 mm				5	
Spring tips 0.5 mm	2**	4**			5
Sprung hook 3.5 mm	1	2			

* pre-fitted to each probe
**1 pre-fitted to each probe, balance supplied separately

 **WARNING**


All accessories are safety-tested. Replace only with Pico Technology accessories.

Maintenance


Changing the probe tip

To change the probe tip use pliers to grip and pull it carefully straight out of its contact socket, along the axis of the probe. Do not grip the white plastic insulator or the housing with pliers, because the tip could be squeezed and made impossible to remove and the probe could be damaged.

After the probe tip is removed, the new tip can be inserted with pliers into the contact socket along the axis of the probe. In order to insert the probe tip completely into the housing, press the probe tip carefully against a hard surface.



Use pliers to grip and pull the probe tip carefully out of its contact socket.



Do not grip the white plastic insulator or the probe housing with pliers.

United Kingdom global headquarters:

Pico Technology
James House
Colmworth Business Park
St. Neots
Cambridgeshire
PE19 8YP
United Kingdom

☎ +44 (0) 1480 396 395

✉ sales@picotech.com

North America regional office:

Pico Technology
320 N Glenwood Blvd
Tyler, TX 75702
United States

☎ +1 800 591 2796

✉ sales@picotech.com

Germany regional office and EU Authorized Representative:

Pico Technology GmbH
Im Rehwinkel 6
30827 Garbsen
Germany

☎ +49 (0) 5131 907 62 90

✉ info.de@picotech.com

Information in this publication supersedes that in all previously published material.

Pico Technology is a registered trade mark of Pico Technology Ltd.

Copyright © 2024 Pico Technology Ltd. All rights reserved.

