

IMPORTANT INFORMATION TO READ and RETURN

Installation Requirements for a Whitley A35 HEPA Workstation

Thank you for choosing one of our products for your laboratory. To enable our engineers to perform an efficient, trouble-free installation please study, complete and fax this form to us on **01274 531197**. Should you have any questions, please do not hesitate to contact us, as we are here to help. When we have received the completed form, our Service Department will contact you to arrange a mutually convenient installation date.

The following information represents the ideal requirement.

Please contact us IMMEDIATELY if your intended location does not match this specification.

					✓			
1	Access Requ	irements						
F	or access, the	e dimensions be	elow should be	taken into account when checking the size of doorways, lifts, stairs, etc.				
5	Space Requi	rements						
The w eight of the equipment is 120kg. If bench mounted, the bench allocated must be flat, level and of sufficient size to support the base fully.								
E	External Dim	ensions*						
F	Width	Depth	Height					
	mm	mm	mm					
	1255	810	710					
*	electrical supp	olies. If benchm	ounted, a furth	allow a localised protrusion of 90mm at rear to accommodate condenser bottle, bracket, gas a ner minimum clearance of 500mm is required above the unit and a minimum clearance of 500 for user/service access.				

Gas Requirements

The incoming gas supplies must be terminated near the right hand side of the main chamber and fitted with leak-proof taps and pressure gauges

The gas lines to which the equipment is attached are the responsibility of the user and should be constructed, tested and maintained to the standards specified within the British Compressed Gasses Association (BCGA) Code of Practice CP4 (or international equivalent). Gas lines previously used for flammable gases must be purged prior to re-use.

Regulators should be fitted in accordance with the information contained in the table below and the various pressures strictly adhered to.

Gas Type	Connection Details	Cylinder Regulator Required	Regulator Outlet Range	Flow Rate	
Anaerobic Gas Mix	1/4" BSP male fitting or connection for 6mm Polyurethane tubing.	Tw o stage	4 - 6 bar (60-90 psi)	40 litres per minute (dynamic)	
	Hydrogen/Anaerobic Gas Mixture Regulator – Tw o Stage – order Code A01745				
Nitrogen	1/4" BSP male fitting or connection for 8mm Polyurethane tubing	Tw o Stage	4 - 6 bar (60-90 psi)	Minimum 150 litres per minute (dynamic)	
	Nitrogen Regulator – Tw o Stage – order Code A01748				

Notes: If the instrument is to be connected to a single cylinder of anaerobic gas mixture only, then this must be capable of delivering 150 litres/minute.

Connection Type							
	Push in connection OR Push on Connection						
Mains Requirements							
Electricity Supply	Wall Socket						
240 volts	1 x Three Pin, 13 Amp. Minimum 6A rating						
Other Considerations							
Do not site the instrument near draughts caused by windows, doors or air conditioning systems							
Remember, if you do not have the required regulators you can order them from Don Whitley Scientific:: • Hydrogen/Anaerobic Gas Mixture Regulator – Tw o Stage – order Code A01745 • Nitrogen Regulator – Tw o Stage – order Code A1748							
Notes							
If an existing unit is being taken in part exchange or is being removed from the laboratory, it must be de-contaminated before DWS staff handle the unit. A certificate or signed letter confirming that the unit has been decontaminated must be given to our engineer.							
In the UK, delivery and installation are free of charge (unless otherwise agreed). If our engineers are unable to install the unit and a return journey is necessary, a charge may be made . Export customers, please refer to your local distributor.							
It is essential that th	is form is completed and returned, to avoid delay to your installatio	n.					
THANK YOU FOR THINKING WHITLEY							

DWS/SSP0005 Form Revision Level: 04

Title

Establishment

Signature

Print Name