

IMPORTANT INFORMATION TO READ and RETURN

Installation Requirements for a Whitley H135 HEPA Hypoxystation

Thank you for choosing one of our products for your laboratory. To enable our engineers to perform an efficient, trouble-free installation please study, complete and fax this form to us on **01274 531197**. Should you have any questions, please do not hesitate to contact us, as we are here to help. When we have received the completed form, our Service Department will contact you to arrange a mutually convenient installation date.

The following information represents the ideal requirement.

Please contact us IMMEDIATELY if your intended location does not match this specification.

| Access Req | uirements | | |
|------------------|--------------------|-----------------------|---|
| Please note th | at this product is | s wider than a | standard door so please ensure that you consider access carefully. |
| Space Requ | irements | | |
| and of sufficier | nt strength to su | pport the unit. | nch mounted, the bench allocated must be flat, level and of sufficient size to support the base fully /orkstation, it must be trolley-mounted. |
| External Din | nensions* | | |
| Width | Depth | Height | |
| mm | mm | mm | |
| 1452 | 1056 | 993 | |
| | nted, a further m | | nce of 500mm is required above the unit and a minimum clearance of 500mm //service access. |
| | | | |

Gas Requirements

The incoming gas supplies must be terminated near the right hand side of the main chamber and fitted with leak-proof taps and pressure gauges.

The gas lines to which the equipment is attached are the responsibility of the user and should be constructed, tested and maintained to the standards specified within the British Compressed Gasses Association (BCGA) Code of Practice CP4 (or international equivalent). Gas lines previously used for flammable gases must be purged prior to re-use.

Regulators should be fitted in accordance with the information contained in the table below and the various pressures strictly adhered to.

| Gas Type | Connection Details | Cylinder Regulator Required | Regulator Outlet Range | Flow Rate |
|-----------------|--|--------------------------------|------------------------|--|
| CO ₂ | 1/4 BSP male fitting or connection for 6mm Polyurethane Tubing CO ₂ Regulator – Two Stage – order Code A01747 | Two stage | 4-6 bar (60-90 psi) | Minimum 10 litres per minute (Dynamic) |
| Air | 1/4 BSP male fitting or connection for 6mm Polyurethane Tubing Air Regulator – Two Stage – order Code – A01554 | Two stage | 4-6 bar (60-90 psi) | Minimum 20 litres per minute (Dynamic) |
| | | | | Continued/ |

| Nitrogen | ¼ BSP male fitting or connection for 8mm Polyurethane Tubing | Two stage | 4-6 bar (60-90 psi) | Minimum 250 litres per minute (Dynamic) |
|--|--|---|--------------------------------|---|
| | Nitrogen Regulator – Two Stage – order Code A01748 | | | |
| | | | | |
| Connection Type | | | | |
| | Push in connec | tion | | |
| | | OK . | | |
| | Push on Conne | ection | | |
| | | | | |
| Mains Requireme | ents | | | |
| Electricity Suppl | | | | |
| 240 volts | 1 x Three Pin, 1 | 3 Amp. Minimum 6A ra | ting | |
| Other Considera | iions | | | |
| Do not site the instru | ment near draughts caused by windov | vs, doors or air condition | ning systems | |
| Carbon DiAir Regula | o not have the required regulators you oxide Regulator – order Code A01747 ttor – Two Stage – order Code A01554 Regulator – Two Stage – order Code A 0 | i. | on Whitley Scientific:: | |
| Notes | | | | |
| | peing taken in part exchange or is bein e unit. | g removed from the lab | oratory, it must be de-conta | minated before |
| unable to inst | elivery and installation are fr all the unit and a return jourr please r ssential that this form is com | ney is necessary, efer to your local | a charge may be r distributor. | nade. Export customers, |
| | THANK YOU F | OR THIN | KING WH | IITLEY |
| | | | | |
| Signature | | Title | | |
| Drint Name | | Fotob | lichment | |