

Declaration of Conformity

No.: OE1-TED-NEP-DEK-148 / 3.0

Name and address of manufacturer

METREL d.d.
Ljubljanska cesta 77
1354 Horjul
Slovenia

Description of the equipment

TeraOhm 5kV MI 3201**Instrument for testing insulation resistance to
5 kVd.c., voltmeter to 600V, microprocessor
controled, mains and battery supply.
CAT IV 600 V**

Reference to the harmonized standards

EN 61010-1:2010 EN 61010-2-030:2010
EN 61010-031:2015
EN 61557-1:2007 EN 61557-2:2007
EN 61326-1:2013 EN 61326-2-2:2013References to the specifications to
which conformity is declared, if
appropriate**LVD Directive 2014/35/EU**
EMC Directive 2014/30/EU
ROHS II Directive 2011/65/EULast two digits of the year in which the
CE marking was affixed**17**

Place and date of issue

Horjul, 2017-02-03

Manufacturer's representative - signature

Edvard Reven



Code:

No part of this document may be reproduced or utilized in any form or by any means without permission in writing from Metrel d.d.