

Measuring and Regulation Equipment Manufacturer

http://www.metrel.si E-mail: metrel@metrel.si METREL d.d. Ljubljanska cesta 77, 1354 HORJUL SLOVENIJA

Phone.: +386 (1) 7558-200 Fax.: +386 (1) 7549-095, 7549-226 Swift Code: LJBASI2X 02923-0018420925 VAT No: SI 56387512

Declaration of Conformity

No.: OE1-TED-NEP-DEK-1695 / 1.0

Name and address of manufacturer	METREL d.d Ljubljanska o 1354 Horjul Slovenia			
Description of the equipment	Digital Transformer Analyser MI 3280			
	Multi-function portable test instrument, for diagnosing of turn ratio, phase deviation and winding resistance of single and three phase transformer. Mains (CAT II 300 V) and battery (Li-ion) supply with IP65 (case closed), IP54 (case opened), microprocessor controlled. Measurement category CAT IV 50 V.			
Reference to the harmonized standards	EN 61010-1:20 EN 61010-031: EN 61010-2-03	2015		326-1:2013 326-2-2:2013
References to the specifications to which conformity is declared, if appropriate	LVD Directive EMC Directive ROHS II Direc			
Last two digits of the year in which the CE marking was affixed	17			
Place and date of issue		Manufacturer	r's repres	sentative - signature
Horjul, 2017-08-22		Edvard Re	even	THS

Code:

No part of this document may be reproduced or utilized in any form or by any means without permission in writing from Metrel d.d.