



DECLARATION OF CONFORMITY

Product Identification:

Fluke 393 - CAT III 1500V TRMS Clamp Meter

First declared: **2021**

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO 9001:2015 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

- Directive 2015/863/EU, UK Regulation 2016: RoHS**
- Directive 2014/30/EU, UK Regulation 2016: Electromagnetic Compatibility (EMC)**
- Directive 2014/35/EU, UK Regulation 2016: Low Voltage (LVD)**

Standards Used:

- EN/BS 63000:2016 Assessment of electrical and electronic products for restriction of hazardous substances
- EN/BS 61326-1:2013 Electrical equipment for measurement, control and laboratory-EMC; General
- EN/BS 61326-2-2:2013 Particular EMC requirements for portable test, measuring and monitoring equipment
- EN/BS 61010-1:2010 Safety requirements for measurement, control and laboratory equipment; General
- EN/BS 61010-031:2015 Safety requirements for hand-held probe assemblies for electrical measurement and test
- EN/BS 61010-2-032:2012 Particular requirements for hand-held and hand manipulated current sensors
- EN/BS 61010-2-033:2012 Particular requirements for hand-held Multimeters for measuring Mains voltage
-

Special Conditions:

None

EU Importer/Distributor

Fluke Europe B.V.
PO Box 1186
5602 BD Eindhoven
The Netherlands

UK Importer/Distributor

Fluke Precision Measurement Ltd.
52 Hurricane Way
Norwich, NR6 6JB
United Kingdom

Manufacturer

Fluke Corporation
6920 Seaway Boulevard M/S 143F
Everett, WA 98203; USA

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation, Everett Washington, USA

DC20210707082951 - This declaration supersedes all previous declarations for this product.

2021-07-07